

# PERFORMANCE COMPARISON OF COMMERCIAL MOLECULAR ASSAYS TO DETECT AND IDENTIFY PANDEMIC INFLUENZA A IN A SENSITIVITY PANEL AND IN CLINICAL SPECIMENS

Astrid Petrich, K. Luinstra, S. Chong, J.B Mahony, and M. Smieja  
St. Joseph's Healthcare, Hamilton, Ontario, Canada

## Revised Abstract

**Objective:** The objective of this study was to compare the sensitivity and specificity of commercially available molecular screening and subtyping assays to detect seasonal and pandemic 2009 H1N1 influenza A.

**Methods:** A panel of 30 samples containing serial dilutions of viral lysate from 3 influenza A subtypes (pandemic 2009 H1N1, seasonal H1N1 and H3N2) and 5 influenza A negative samples was tested with 4 commercial influenza A screening and subtyping assays. The panel was designed to challenge the lower limit of detection in the 4 assays; the Influenza Screen & Type RT-PCR Kit (astra Diagnostics), Influenza A/H1N1 Detection Set (Roche), Lumindex x-Tag™ RVP fast and ProFLU™/ProFLU™-ST assays (Prodesse) were tested according to manufacturer's instructions. The performance of these assays was further investigated using 100 NPS samples which included 51 influenza A negatives and 10 seasonal H1N1, 10 seasonal H3N2 and 29 pandemic H1N1 positive samples as previously determined by an influenza A matrix PCR. The sensitivity and specificity of each assay was determined using a combined reference standard of positivity in at least 2 assays.

**Results:** The commercial assays were able to detect 27 to 60 % of the 15 potential pandemic H1N1 positives and 50 to 80 % of the 10 potential seasonal flu positives in the panel. For the clinical samples, the commercial assays had a sensitivity and specificity for influenza A screening ranging from 82 to 100%. The pandemic H1N1 subtyping assays had a sensitivity and specificity ranging from 92 to 100%. With the clinical samples, only the Influenza Screen and Type RT-PCR assay from astra Diagnostics allowed accurate simultaneous identification of influenza A and 2009 H1N1 (sensitivity 100% specificity 100%).

**Conclusions:** The performance of commercial assays for influenza A screening, particularly for pandemic H1N1, was variable. Proper validation of the screening assay is therefore critical for the identification of pandemic H1N1.

## Objective

The objective of this study was to compare the sensitivity and specificity of commercially available molecular screening and subtyping assays to detect seasonal and pandemic 2009 H1N1 influenza A.

## Methods

**Commercial assays:**

- Influenza Screen & Type RT-PCR Kit (astra Diagnostics)
- Influenza A/H1N1 Detection Set (Roche): Ver 1 (05/ 2009), Ver 2 (08/2009)
- x-Tag™ RVP fast (Luminex)
- ProFLU™+ /ProFLU™-ST assays (Prodesse)

Assays were tested according to manufacturer's instructions

**Sensitivity panel:**

Panel of extracted nucleic acid from 30 randomized samples consisting of viral lysate dilutions of a pandemic isolate (A/swine/California/04/2009-like), a seasonal H3N2 strain (A/Victoria/3/75), a seasonal H1N1 strain (A/Solomon Islands/03/06) and 5 influenza A negative samples (Flu B, hMPV, Para 3 and 2 water blanks)

**Clinical Evaluation:**

- 100 NPS samples (Copan FLOQSwabs™) collected in UTM-RT (Copan): 49 influenza A positives (29 pandemic H1N1, 10 seasonal H1N1, 10 seasonal H3N2) and 51 influenza A negative samples as previously determined by an influenza A matrix PCR
- Purified nucleic acid from 200 uL of NPS sample extracted by easyMAG (bioMérieux), eluted in 60 uL and tested in the 4 commercial assays
- The sensitivity and specificity of each assay was determined using a combined reference standard of positivity in at least 2 assays

## Results of Sensitivity Panel

Sensitivity Panel		astra Diagnostics	Roche Diagnostics	Luminex	Prodesse
#	Panel Contents	Influenza A Screen & Type RT-PCR	2009 H1N1 Detection Set Ver 1	x-Tag™ RVP fast	ProFLU™ plus & ProFLU™-ST
1	2009 H1N1 Dilution 1-1	15	2009 H1N1	Flu A	2009 H1N1
2	2009 H1N1 Dilution 1-2	28	2009 H1N1	Flu A	2009 H1N1
3	2009 H1N1 Dilution 1-3	24	2009 H1N1	Flu A	2009 H1N1
4	2009 H1N1 Dilution 2-1	7	2009 H1N1	Flu A	2009 H1N1
5	2009 H1N1 Dilution 2-2	23	2009 H1N1	Flu A	Flu A; untypeable
6	2009 H1N1 Dilution 2-3	30	2009 H1N1	Flu A	2009 H1N1
7	2009 H1N1 Dilution 3-1	10	2009 H1N1	Neg	Flu A; untypeable
8	2009 H1N1 Dilution 3-2	27	2009 H1N1	Neg	Flu A; untypeable
9	2009 H1N1 Dilution 3-3	16	2009 H1N1	Neg	Neg
10	2009 H1N1 Dilution 4-1	20	Neg	Negative	Flu A; untypeable
11	2009 H1N1 Dilution 4-2	11	Neg	Negative	Neg
12	2009 H1N1 Dilution 4-3	3	Neg	Negative	Neg
13	2009 H1N1 Dilution 5-1	22	Neg	Negative	Neg
14	2009 H1N1 Dilution 5-2	5	Neg	Negative	Neg
15	2009 H1N1 Dilution 5-3	19	Neg	Negative	Neg
16	H1N1 Seasonal Dilution 1	48	Flu A	Flu A, H1	Flu A; H1
17	H1N1 Seasonal Dilution 2	4	Flu A	Flu A, H1	Flu A; H1
18	H1N1 Seasonal Dilution 3-1	1	Neg	Flu A	Flu A; H1
19	H1N1 Seasonal Dilution 3-2	14	Neg	Negative	Flu A; untypeable
20	H1N1 Seasonal Dilution 3-3	6	Flu A	Negative	Flu A; H1
21	H3N2 Seasonal Dilution 1	29	Flu A	Flu A, H3	Flu A; H3
22	H3N2 Seasonal Dilution 2	26	Flu A	Flu A, H3	Flu A; untypeable
23	H3N2 Seasonal Dilution 3-1	13	Flu A	Neg	Flu A; untypeable
24	H3N2 Seasonal Dilution 3-2	8	Flu A	Neg	Flu A; untypeable
25	H3N2 Seasonal Dilution 3-3	17	Flu A	Neg	Flu A; untypeable
26	dh20	21	Neg	Negative	Neg
27	dh20	25	Neg	Negative	Neg
28	Flu B	2	Neg	Neg	Flu B
29	hMPV	9	Neg	Neg	MPV
30	Para 3	12	Neg	Neg	Para 3

### Panel Sensitivity for 2009 H1N1

astra Diagnostics

• 9/15 = 60.0%

Roche Diagnostics

• 6/15 = 26.7 %; 2 untypeable influenza A +ve

Luminex

• 7/15 = 46.7 %; 1 equivocal influenza A +ve

Prodesse

• 5/15 = 33.3 %; 4 untypeable influenza A +ve

### Panel Sensitivity for Seasonal Influenza

astra Diagnostics

• 8/10 = 80.0%; not typed

Roche Diagnostics

• 5/10 = 50.0%; not typed

Luminex

• 6/10 = 60.0 %; 1 equivocal

• H1, 2/5 = 40.0%; H3, 4/5 = 80.0%

Prodesse

• 5/10 = 50.0%; 4 untypeable

• H1, 4/5 = 80.0%; H3, 1/5 = 20.0%

## Clinical Performance

astra Diagnostics Influenza Screen & Type RT-PCR Kit:

- no false negatives or false positives

Roche Diagnostics Influenza A/H1N1 Detection Set:

- InfA/M2 assay:

- Ver 1, 4 pandemic H1N1 false negatives
- Ver 1, 1 seasonal H1N1 false negative
- Ver 2, 2 false positive influenza A negative samples

InfA/H1 (swine) assay: no false negatives or false positives

Luminex x-Tag™ RVP fast:

- 3 pandemic H1N1 false negatives

Prodesse ProFLU™+ assay:

- 1 pandemic H1N1 false negative

Prodesse ProFLU-ST assay:

- 4 pandemic H1N1 false negatives

### Clinical Sensitivity and Specificity

	astra Screen & Type	Roche inf A/M2 Ver 1	Roche inf A/M2 Ver 2	Roche inf A/H1 Ver 1	Roche inf A/H1 Ver 2	Luminex x-Tag™ RVP fast	Prodesse ProFLU™+	Prodesse ProFLU™-ST
<b>Sensitivity</b>	49/49 (100%)	23/28 (82.1%)	21/21 (100%)	17/17 (100%)	12/12 (100%)	46/49 (93.9%)	48/49 (98.0%)	45/49 (91.8%)
<b>Specificity</b>	51/51 (100%)	22/22 (100%)	27/29 (93.1%)	33/33 (100%)	38/38 (100%)	51/51 (100%)	51/51 (100%)	51/51 (100%)
<b>PPV</b>	49/49 (100%)	23/23 (100%)	21/23 (91.3%)	17/17 (100%)	12/12 (100%)	46/46 (100%)	48/48 (100%)	45/45 (100%)
<b>NPV</b>	51/51 (100%)	22/27 (81.5%)	27/27 (100%)	33/33 (100%)	38/38 (100%)	51/54 (94.4%)	51/52 (98.1%)	51/55 (92.7%)

## Results of Clinical Evaluation

Sample Breakdown	astra Diagnostics	Roche Diagnostics Ver 1		Roche Diagnostics Ver 2		Luminex	Prodesse	
	Screen & Type	InfA/M2	InfA/H1 (swine)	InfA/M2	InfA/H1 (swine)	x-Tag™ RVP fast	ProFLU™ plus	ProFLU™-ST
29 2009 H1N1+	29/29 2009 H1N1+	13/17 Flu A+	17/17 2009 H1N1+	12/12 Flu A+	12/12 2009 H1N1+	25/29 Flu A+	28/29 Flu A+	25/29 2009 H1N1+
10 H1N1+	20/20 Flu A+	10/11 Flu A+	11/11 2009 H1N1-	9/9 Flu A+	9/9 2009 H1N1-	10/10 H1+	20/20 Flu A+	10/10 H1+
10 H3N2+						10/10 H3+		10/10 H3+
51 Flu A-	51/51 Flu A-	22/22 Flu A-	22/22 2009 H1N1-	27/29 Flu A-	29/29 2009 H1N1-	51/51 Flu A-	51/51 Flu A-	51/51 Flu A-
100 Total	100 Tested	50 Tested	50 Tested	50 Tested	50 Tested	100 Tested	100 Tested	100 Tested

- The 51 influenza A negative samples included 11 samples positive for other respiratory viruses: 1 adenovirus, 3 RSV- A, 4 enterovirus/rhinovirus, 1 influenza B, 1 para 1 and 1 para 4

## Conclusions

- The sensitivity of the commercial assays for influenza A screening, particularly for pandemic H1N1, was variable (82.1-100%)
- Proper validation of an influenza A screening assay is critical to the identification of 2009 H1N1
- Only the Influenza Screen and Type RT-PCR assay from astra Diagnostics allowed simultaneous identification of influenza A and 2009 H1N1 with a 100% sensitivity and specificity, in clinical samples